



Engineering and Testing for EMC and Safety Compliance

CONFORMANCE STATEMENT

APPLICABLE DIRECTIVES

EMC Directive: Council Directive 93/42/EEC: June 1993 Concerning Medical Devices (12.7.93 Jo No L 169/1)

Standards And Specific Tests To Which Conformity Is Declared

STANDARDS	ENVIRONMENTAL PHENOMENA	APPLICABILITY
		<input checked="" type="checkbox"/> Tested <input type="checkbox"/> Not Tested
EN 60601-1-2: 2002	Emissions (EN55011: 1998)	<input type="checkbox"/>
EN 60601-1-2: 2002	Harmonic Current Emissions (EN61000-3-2: 1995, A1, and A2: 1998, A14: 2000)	<input type="checkbox"/>
EN 60601-1-2: 2002	Voltage Fluctuations and Flicker (EN61000-3-3: 1995)	<input type="checkbox"/>
EN 60601-1-2: 2002	Electrostatic Discharge (EN61000-4-2: 1995, A1: 1998, A2: 2001)	<input checked="" type="checkbox"/>
EN 60601-1-2: 2002	Radiated Immunity (EN 61000-4-3: 2002, A1: 2002)	<input checked="" type="checkbox"/>
EN 60601-1-2: 2002	Electrical Fast Transients (EN61000-4-4: 1995, A1 and A2: 2001)	<input checked="" type="checkbox"/>
EN 60601-1-2: 2002	Surge (EN61000-4-5: 1995, A1: 2001)	<input type="checkbox"/>
EN 60601-1-2: 2002	Conducted Immunity (EN61000-4-6: 1996, A1: 2001)	<input type="checkbox"/>
EN 60601-1-2: 2002	Magnetic Immunity (EN61000-4-8: 1993)	<input checked="" type="checkbox"/>
EN 60601-1-2: 2002	Voltage Dips and Short Interruptions (EN61000-4-11; 1994, A1: 2001)	<input type="checkbox"/>

Equipment Under Test

Manufacturer's Name	Meru Networks, Inc.
Manufacturer's Address, Phone and Fax Numbers	1309 S. Mary Ave. Sunnyvale, CA 94087 408-215-5300
Type of Equipment	Medical Electrical
Equipment Classification	Wireless Network Access Point
Models	AP150, AP200
Year of Manufacture	2006

We, the undersigned, hereby declare that the equipment specified above conforms to the above-identified standard as described in the attached test records.

Test Engineer:

Jon Wilson

Signature

July 11, 2006

Date

Supervising Engineer
& RTL President:

Desmond A. Fraser

Signature

July 11, 2006

Date